

# KIMTECH™ STERLING™ Nitrile Exam Glove, 9.5

**Product Information**

Code	Size	Palm Width*(mm)	Length* (mm)	Quantity
50705	XS	70	242	200 gloves/box; 10 boxes/case
50706	S	80	242	200 gloves/box; 10 boxes/case
50707	M	95	242	200 gloves/box; 10 boxes/case
50708	L	110	242	200 gloves/box; 10 boxes/case
50709	XL	120	242	170 gloves/box; 10 boxes/case

\*As measured by ASTM D 6319; 2.5 AQL<sup>1</sup>



Property	Description
General Product Information	For use in Food Handling and Food Processing; Cleared for use in chemotherapy; meets USP 800 guidelines; tested for use 16 with hazardous drugs – 12 chemotherapy drugs and 4 hormones per ASTM D6978-05
General Description	Light Gray nitrile, powder free, exam, non-sterile, textured fingertips, ambidextrous gloves, beaded cuff, both sides chlorinated, 9.5"

Physical Property	Unit	Upper Limit	Test Method
Watertightness	NA	Pass	ASTM D 5151 (1.0 AQL)
Residual Powder	mg/glove	2	ASTM D 6124 (2.5 AQL)

Thickness	Target	Target	Test Method (2.5 AQL)
Middle Finger	3.54 mil	0.09 mm	ASTM D 6319
Palm	3.15 mil	0.08 mm	ASTM D 6319
Cuff	2.36 mil	0.06 mm	ASTM D 6319

Property	Tensile Strength	Ultimate Elongation
Before Aging	42.0 Mpa	650%
After Aging	38.0 Mpa	550%
Testing Method (2.5 AQL)	ASTM D 412 ASTM D 573 ASTM D 6319	ASTM D 412 ASTM D 573 ASTM D 6319

<sup>1</sup> AQL levels defined according to ISO 2859-1 (current edition) Sampling Procedures for Inspection by Attributes.

**Workmanship Requirement:** Gloves shall be made in accordance with reasonable industry practice with respect to defects, dirt and contamination. Gloves should be packaged as specified having described size and color in the appropriate quantity as listed. Shipping Case and inner packaging shall be labeled with product name, size, lot/batch and catalog number.

For Additional information: Contact your Kimberly-Clark representative, our Customer Service Team (888)-346-GOKC (4652) or [KCPIInfo@kcc.com](mailto:KCPIInfo@kcc.com)